



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

May 3, 2013

Little Doctor Electronics (Nantong) Co., Ltd.
c/o Ms. Selina Qian
No. 8, Tongxing Road,
Nantong Economic & Technological Development Area
Nantong, Jiangsu
CHINA 226009

Re: K123600

Trade/Device Names: Automatic Digital Blood Pressure Monitor, Model LD-568,
LD-520, LD-522, LD-587 & LD-528

Regulatory Number: 21 CFR 870.1130

Regulation Name: Noninvasive Blood Pressure Measurement System

Regulatory Class: Class II (Two)

Product Code: DXN

Dated: March 19, 2013

Received: April 3, 2013

Dear Ms. Qian:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Ms. Selina Qian

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Owen P. Faris -S

for Bram D. Zuckerman
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

INDICATIONS FOR USE

Page 1 of 1

510(k) Number (if known): K123600

Device Name: **Electronic Blood Pressure Monitor**

Model: LD-568, LD-520, LD-522, LD-587 & LD-528

Indication for Use:

The Electronic Blood Pressure Monitor LD-568, LD-520, LD-522, LD-587 & LD-528 is device intended to measure the systolic and diastolic blood pressure and pulse rate of an adult individual by using a non-invasive technique in which an inflatable cuff is wrapped around the arm.

Prescription Use _____

(Per 21 CFR 801 Subpart D)


Over-The Counter Use X

(21 CFR 807 Subpart C)

OR

(PLEASE DO NOT WRITE BELOW THIS LINE -CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



Owen P. Faris -S
2013.05.03 12:52:51
04'00"